

Issued by	NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 9 of Directive 2004/22/EC, after having established that the Measuring instrument meets the applicable requirements of Directive 2004/22/EC, to:
Manufacturer	Gilbarco GmbH Ferdinand-Henze-Straße 9 D-33154 Salzkotten Germany
Measuring instrument	A Fuel Dispenser Type : 397; ARAL Accuracy class : 0,5 Environment classes : M2 / E1 Temperature range liquid : -40 °C / +50 °C Temperature range ambient : -25 °C / +55 °C -40 °C / +55 °C for electronics, in case a heater is applied $Q_{\min} - Q_{\max}$: See the Description §1.2. Further properties are described in: - Description T10057 revision 3; - Documentation folder T10057-1.
Valid until	5 April 2017
Remarks	- Revision 3 was issued due to the addition of several self-service devices. This revision replaces the earlier version, except for its documentation folder; - The fuel dispenser may be combined with one or more LPG- and/or gasoline oil dispensers and/or measuring installations for AdBlue, which make use of the same or separate calculating/indicating device(s).

Issuing Authority

NMi Certin B.V., Notified Body number 0122
22 February 2010


C. Oosterman
Head of the Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The designation of NMi Certin BV, as Notified Body can be verified at <http://ec.europa.eu/enterprise/newapproach/nando/>

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).

Reproduction of the complete document only is permitted.