



Issued by NMI Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands

Notified Body number 122

In accordance with The "Metrologiewet" (Stb. 2006, 137) as Dutch implementation of Directive 2004/22/EC on measuring instruments (MID).

Applicant ODS B.V.
Donk 6
2992 LE BARENDRECHT
The Netherlands

In respect of **A measuring system for oil**

Manufacturer : ODS B.V.
Type : Rheonik ODS Measuring System

Characteristics $Q_{min} - Q_{max}$: 400 – 2050 t/h
Accuracy class : 0,3
Environment classes : M2 / E2
Temperature range liquid : +35 °C / +45 °C
Temperature range ambient : -25 °C / +55 °C
Maximum pressure : 30 bar(g)
Intended for the measurement of : Oil

In the description number T10139 revision 0 further characteristics are described.

Valid until 23 June 2018.

Description and documentation The measurement installation is described in the description T10139 revision 0 and documented in the documentation folder T10139-1, appertaining to this EC-type examination certificate.

Dordrecht, 23 June 2008.
NMI Certin B.V.

Ing. C. Oosterman
Manager Product Certification