

EC-type examination certificate

Number **T10178** revision **0** Project number: 804990 Page 1 of 1

Issued by

NMi Certin B.V.

Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands

In accordance with

The "Metrologiewet" (Stb 2006, 137) as Dutch implementation of the Directive

2004/22/EC on measuring instruments (MID)

Applicant

FMC Kongsberg Metering AS

Kirkegårdsveien 45 N-3116 Kongsberg

Norway

Submitted

A non interruptible measuring instrument on a skid.

Manufacturer : FMC Kongsberg

Designation : Fiscal Oil Metering System

Characteristics

Qmin - Qmax

: See paragraph 1.2 of the description

Minimum measured quantity

See paragraph 1.2 of the description

Accuracy class

0.3

Environment class

M1 / E2 -10 °C / +50 °C

Temperature range liquid Temperature range ambient

-25 °C / +55 °C; condensing humidity

open location

Intended for the measurement of

Oil and oil products.

In the description T10178 revision 0 the additional characteristics are given.

Valid until

6 April 2019.

Description and Documentation

The measuring instrument is described in the description number T10178 revision 0 and documented in the documentation folder number T10178-1,

appertaining to this EC-type examination certificate.

The Notified Body no. 0122 NMi Certin, 6 April 2009.

C. Oosterman

Head Certification Board

NMi Certin BV Hugo de Grootplein 1 3314 EG Dordrecht PO Box 394 3300 AJ Dordrecht, NL T +31 78 6332 332 F +31 78 6332 309 certin@nmi.nl www.nmi.nl Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see "Regulation objection and appeal against decisions of NMi")

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

Reproduction of the complete document is

