

Issued by NMI Certin B.V.  
Hugo de Grootplein 1  
3314 EG Dordrecht  
The Netherlands

In accordance with The "Metrologiewet" (Stb. 2006, 137) as Dutch implementation of Directive 2004/22/EC on measuring instruments (MID).

Applicant Brand Bierbrouwerij B.V.  
Brouwerijstraat 2-10  
6321 AG Wijlre  
The Netherlands

In respect of **A measuring system for liquids other than water**

Manufacturer : Brand Bier Brouwerij B.V.  
Type : Barrel loading System: Multimat

Characteristics  $Q_{\min} - Q_{\max}$  : 20 – 200 L/min  
Accuracy class : 0,5  
Environment classes : M2 / E2  
Temperature range liquid : -5 °C / + 35 °C  
Temperature range ambient : -10 °C / +40 °C  
Maximum pressure : 10 bar(g)  
Intended for the measurement of : Beer

In the description number T10183 revision 0 further characteristics are described.

Valid until 12 January 2018.

Description and documentation The measurement installation is described in the description T10183 revision 0 and documented in the documentation folder T10183-1, appertaining to this EC-type examination certificate.

Notified Body number 0122  
NMI Certi, 27 March 2009



C. Oosterman  
Head Certification Board