

Issued by	NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 9 of Directive 2004/22/EC, after having established that the Measuring instrument meets the applicable requirements of Directive 2004/22/EC, to:
Applicant	FMC Kongsberg Metering AS Kirkegårdsveien 45 N-3116 Kongsberg Norway
Submitted	A non-interruptible measuring instrument on a skid. Manufacturer : FMC Kongsberg Designation : Fiscal Metering System
Characteristics	Qmin – Qmax : See paragraph 1.2 of the description Minimum measured quantity : See paragraph 1.2 of the description Accuracy class : 1.0 (N-butane and I-Butane); 1.5 (Propane) Environment class : M1 / E2 Temperature range liquid : -50 °C / +50 °C Temperature range ambient : -25 °C / +55 °C; condensing humidity; open location Intended for the measurement of : Liquefied Propane, liquefied N-Butane and liquefied I-Butane. In the description T10254 revision 0 the additional characteristics are given.
Valid until	3 February 2022.
Description and Documentation	The measuring instrument is described in the description number T10254 revision 0 and documented in the documentation folder number T10254-1, appertaining to this EC type-examination certificate.

Issuing Authority

NMi Certin B.V.
3 February 2012


C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The designation of NMi Certin BV.as Notified Body can be verified at <http://ec.europa.eu/enterprise/newapproach/nando>

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).

Reproduction of the complete document only is permitted.