


Issued by	NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 9 of Directive 2004/22/EC, after having established that the Measuring instrument meets the applicable requirements of Directive 2004/22/EC, to:
Manufacturer	Faure Herman Route de Bonnétable 72400 La Ferté Bernard France
Measuring instrument	An interruptible measuring system for liquids Type : Heliflu - ML Destined for the measurement of : Hydrocarbons, alcohol, potable liquids and liquid chemicals Accuracy class : 0,5 Environment classes : M2 / E2 Temperature range liquid : -30 °C / +180 °C Temperature range ambient : -25 °C / +55 °C Q _{min} - Q _{max} : See § 1.2 of the description Minimum measured quantity : See § 1.2 of the description Further properties are described in the annexes: – Description T10286 revision 0; – Documentation folder T10286-1.
Valid until	11 June 2020

Issuing Authority **NMi Certin B.V., Notified Body number 0122**
11 June 2010


C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The designation of NMi Certin BV.as Notified Body can be verified at <http://ec.europa.eu/enterprise/newapproach/nando/>

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).

Reproduction of the complete document only is permitted.