

Issued by

NMi Certin B.V.,  
designated and notified by the Netherlands to perform tasks with respect to  
conformity assessment procedures mentioned in article 17 of Directive  
2014/32/EU, after having established that the Measuring instrument meets  
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer

Endress+Hauser Italia S.p.A.  
Via Fratelli di Dio, 7  
20063 Cernusco sul Naviglio, Milano  
Italy

Measuring instrument

An interruptible or non-interruptible **measuring system** for liquids other  
than water, based on a dynamic mass (Coriolis) meter.

Manufacturer's mark or name : Endress + Hauser

Type designation : MID 001

Accuracy class : 0,3; 0,5; 1,0; 1,5 or 2,5

Destined for the measurement of : Oil and oil products, chemicals,  
potable liquids and liquefied gasses  
under pressure with densities  
between 400 and 1400 kg/m<sup>3</sup>.

Further properties are described in the annexes:

- Description T10288 revision 7;
- Documentation folder T10288-2.

Valid until

11 June 2030

Remarks

- This revision replaces the previous revisions.
- The documentation folder is not changed.
- The measuring system is approved for measuring mass, density and  
volume.

Issuing Authority

**NMi Certin B.V., Notified Body number 0122**  
24 November 2020

Certification Board

**NMi Certin B.V.**  
Thijsseweg 11  
2629 JA Delft  
The Netherlands  
T +31 88 636 2332  
[certin@nmi.nl](mailto:certin@nmi.nl)  
[www.nmi.nl](http://www.nmi.nl)

This document is issued under the provision that  
no liability is accepted and that the  
manufacturer shall indemnify third-party  
liability.

The designation of NMi Certin B.V. as Notified  
Body can be verified at  
[http://ec.europa.eu/growth/tools-  
databases/nando/](http://ec.europa.eu/growth/tools-databases/nando/)

Reproduction of the complete  
document only is permitted.

This document is digitally  
signed and sealed. The digital  
signature can be verified in the  
blue ribbon at the top of the  
electronic version of this  
certificate.