

EC type-examination Certificate

Number **T10302** revision 2
Project number 12200541
Page 1 of 1

Issued by	NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 9 of Directive 2004/22/EC, after having established that the Measuring instrument meets the applicable requirements of Directive 2004/22/EC, to:
Manufacturer	Red Seal Measurement 1310 Emerald Rd Greenwood, SC 29646 United States of America
Measuring instrument	An interruptible measuring system on a road tanker Brand : Red Seal Measurement or Itron Type : RML2000 $Q_{min} - Q_{max}$: 36 - 272 kg/min Accuracy class : 1,0 Environment classes : M3 / E3 Temperature range liquid : -10 °C / +50 °C Temperature range ambient : -25 °C / +55°C Maximum Pressure : 24 bar(g) Minimum measured quantity : 50 kg Destined for the measurement of : LPG Further properties are described in the annexes: – Description T10302 revision 2; – Documentation folder T10302-1.
Valid until	21 December 2020
Remarks	This revision replaces the earlier versions, except for its documentation folder.

Issuing Authority **NMi Certin B.V., Notified Body number 0122**
18 December 2012

C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision
that no liability is accepted and that the
applicant shall indemnify third-party liability.

The designation of NMi Certin BV.as Notified
Body can be verified at [http://
ec.europa.eu/enterprise/newapproach/nando/](http://ec.europa.eu/enterprise/newapproach/nando/)

Parties concerned can lodge
objection against this decision,
within six weeks after the date of
submission, to the general manager
of NMi (see www.nmi.nl).

Reproduction of the complete
document only is permitted.