

Issued by	NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 9 of Directive 2004/22/EC, after having established that the Measuring instrument meets the applicable requirements of Directive 2004/22/EC, to:
Manufacturer	Egemin Automation Baarbeek 1 B-2070 Zwijndrecht Belgium
Measuring instrument	An interruptible measuring system Type : ATPC – BPO jet fuel injectie Destined for the measurement of : Oil and oil products Accuracy class : 0,3 Environment classes : M2 / E2 Temperature range liquid : -10 °C / +140 °C Temperature range ambient : -25 °C / +55 °C $Q_{min} - Q_{max}$: 100 – 520 m ³ /h Minimum measured quantity : See § 1.2 of the description Further properties are described in the annexes: – Description T10342 revision 0; – Documentation folder T10342-1.
Valid until	28 March 2023
Issuing Authority	NMi Certin B.V., Notified Body number 0122 28 March 2013

C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision
that no liability is accepted and that the
applicant shall indemnify third-party liability.

The designation of NMi Certin BV.as Notified
Body can be verified at [http://
ec.europa.eu/enterprise/newapproach/nando/](http://ec.europa.eu/enterprise/newapproach/nando/)

Parties concerned can lodge
objection against this decision,
within six weeks after the date of
submission, to the general manager
of NMi (see www.nmi.nl).

Reproduction of the complete
document only is permitted.