

Issued by	NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 9 of Directive 2004/22/EC, after having established that the Measuring instrument meets the applicable requirements of Directive 2004/22/EC, to:
Manufacturer	Total Belgium Zoning Industriel, Zone A B-7181 Feluy Belgium
Measuring instrument	An interruptible measuring system for liquids other than water. Type : TOTAL-MID 001 Destined for the measurement of : Alcohol Accuracy class : 0,5 Environment classes : M3 / E3 Temperature range liquid : See § 1.2 of the description Temperature range ambient : See § 1.2 of the description $Q_{\min} - Q_{\max}$: See § 1.2 of the description Minimum measured quantity : See § 1.2 of the description Further properties are described in the annexes: – Description T10344 revision 1; – Documentation folder T10344-1.
Valid until	11 February 2021
Remarks	The measuring instrument is approved for measuring mass, density at metering conditions, density at base conditions, volume at metering conditions and volume at base conditions. This revision replaces the earlier version, except for its documentation folder and is issued due to a change of the manufacturer.

Issuing Authority **NMi Certin B.V., Notified Body number 0122**
23 February 2011


C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.
The designation of NMi Certin BV.as Notified Body can be verified at <http://ec.europa.eu/enterprise/newapproach/nando/>

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMI (see www.nmi.nl).

Reproduction of the complete document only is permitted.