

Issued by	NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 9 of Directive 2004/22/EC, after having established that the Measuring instrument meets the applicable requirements of Directive 2004/22/EC, to:
Manufacturer	Amtra Engineering bv Pieter Zeemanweg 97 3316 GZ Dordrecht The Netherlands
Measuring instrument	An interruptible measuring system on a skid Type : Flex Fill F100 Destined for the measurement of : Liquid petroleum and related products, potable liquids, liquid chemicals and liquefied gasses under pressure with densities between 450 and 1100 kg/m ³ Accuracy class : 0.5 Environment classes : M3 / E3 Temperature range liquid : -10 °C / 50 °C Temperature range ambient : -25 °C / 55 °C Q _{min} - Q _{max} : 18 - 272 kg/min Minimum measured quantity : See § 1.2 of the description Further properties are described in the annexes: – Description T10408 revision 0; – Documentation folder T10408-1.
Valid until	11 June 2022
Remarks	The system is only approved for mass delivery.
Issuing Authority	NMi Certin B.V., Notified Body number 0122 11 June 2012

C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The designation of NMi Certin BV.as Notified Body can be verified at <http://ec.europa.eu/enterprise/newapproach/nando/>

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).

Reproduction of the complete document only is permitted.