

Issued by	NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 9 of Directive 2004/22/EC, after having established that the Measuring instrument meets the applicable requirements of Directive 2004/22/EC, to:
Manufacturer	VIS Elektrotechniek B.V. A.Banckerstraat 1c 3115 JE Schiedam The Netherlands
Measuring instrument	An interruptible measuring system on a road tanker (truck) Type : VISEL 1 Destined for the measurement of : AdBlue (Aqueous liquid) Accuracy class : 0,5 Environment classes : M3 / E3 Temperature range liquid : -10 °C / +50 °C Temperature range ambient : -25 °C / +55 °C $Q_{min} - Q_{max}$: 50 L/min- 600 L/min Minimum measured quantity : See § 1.2 of the description Further properties are described in : – Description T10519 revision 0; – Documentation folder T10519-1.
Valid until	22 November 2022

Issuing Authority **NMi Certin B.V., Notified Body number 0122**
22 November 2012

C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The designation of NMi Certin BV.as Notified Body can be verified at <http://ec.europa.eu/enterprise/newapproach/nando/>

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).

Reproduction of the complete document only is permitted.