

Issued by	NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 9 of Directive 2004/22/EC, after having established that the Measuring instrument meets the applicable requirements of Directive 2004/22/EC, to:
Manufacturer	Emerson Process Management Romania SRL 22, Zlaten Rog Street 1407 Sofia Bulgaria
Measuring instrument	An interruptible or non-interruptible LPG measuring system Manufacturer : Emerson Type : EPM-MMI-MID005-BG-001 Minimum – maximum flow rate : 57 – 9000 kg/min ($Q_{min} - Q_{max}$) Minimum measured quantity : See § 1.2 of the description Maximum pressure : 16 bar(g) Accuracy class : 1.0 Environment classes : M2 / E2 Temperature range liquid : 0 °C / +40 °C Temperature range ambient : -25 °C / +55 °C Intended for the measurement of : Liquefied propane under pressure, liquefied butane under pressure and mixtures thereof (LPG).
	Further properties are described in: – Description T10656 revision 0; – Documentation folder T10656-1.
Valid until	22 October 2024
Remarks	– The measuring system is approved for measuring mass only.

Issuing Authority

NMi Certin B.V., Notified Body number 0122
22 October 2014


C. Oosterman
Head Certification Board

NMi Certin BV
Hugo de Grootplein 1
3314 EG Dordrecht
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.
The designation of NMi Certin B.V. as Notified Body can be verified at <http://ec.europa.eu/enterprise/newapproach/nando>

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).

Reproduction of the complete document is permitted.