


Issued by	NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 9 of Directive 2004/22/EC, after having established that the Measuring instrument meets the applicable requirements of Directive 2004/22/EC, to:
Manufacturer	Emerson Process Management Logie Court, Stirling University – Innovation Park FK9 4N, Stirling Scotland
Measuring instrument	An interruptible <b>measuring system</b> Type : LPG Fiscal Metering Destined for the measurement of : LPG Accuracy class : 1.0 Environment classes : M1 / E2 Temperature range liquid : -30 °C / +50 °C Temperature range ambient : -25 °C / +55 °C $Q_{\min} - Q_{\max}$ : 2590 – 36553 kg/hr Minimum measured quantity : See § 1.2 of the description  Further properties are described in the annexes: – Description T10737 revision 0; – Documentation folder T10737-1.
Valid until	22 January 2026

Issuing Authority **NMi Certin B.V., Notified Body number 0122**  
22 January 2016



C. Oosterman  
Head Certification Board

**NMi Certin B.V.**  
Hugo de Grootplein 1  
3314 EG Dordrecht  
The Netherlands  
T +31 78 6332332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision  
that no liability is accepted and that the  
manufacturer shall indemnify third-party  
liability.

The designation of NMi Certin BV.as Notified  
Body can be verified at [http://  
ec.europa.eu/enterprise/newapproach/nando/](http://ec.europa.eu/enterprise/newapproach/nando/)

Parties concerned can lodge  
objection against this decision,  
within six weeks after the date of  
submission, to the general manager  
of NMi (see [www.nmi.nl](http://www.nmi.nl)).

Reproduction of the complete  
document only is permitted.