



EU-type examination certificate

Number **T10773** revision 0
Project number 16200228
Page 1 of 1

Issued by	NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:
Manufacturer	Emerson Process Management Limited - Daniel Measurement And Control Logie Court, Stirling University Innovation Park Stirling Scotland
Measuring instrument	An interruptible measuring system for liquids other than water Type : Coriolis metering skid Tag number – Station no. 2 Destined for the measurement of : Black product (viscosity around 380 cSt) Accuracy class : 0,3 Environment classes : M1 / E2 Temperature range liquid : +5 °C / +50 °C Temperature range ambient : -10 °C / +55 °C $Q_{min} - Q_{max}$: 332 – 19960 kg/min Minimum measured quantity : See § 1.2 of the description Further properties are described in the annexes: – Description T10773 revision 0; – Documentation folder T10773-1.
Valid until	28 December 2026

Issuing Authority **NMi Certin B.V., Notified Body number 0122**
28 December 2016


C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMi Certin BV.as Notified Body can be verified at <http://ec.europa.eu/>

Reproduction of the complete document only is permitted.

