

Issued by	NMI Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:	
Manufacturer	Budelse Brouwerij B.V. Nieuwstraat 9 6021 HP Budel The Netherlands	
Measuring instrument	An interruptible <b>measuring system</b>	
Type	:	IZM-TEVG1
Manufacturer's mark or name	:	Budelse Brouwerij
Destined for the measurement of	:	Beer
Accuracy class	:	0,5
Environment classes	:	M3 / E3
Temperature range liquid	:	-5 °C / +35 °C
Temperature range ambient	:	-25 °C / +55 °C
$Q_{min} - Q_{max}$	:	20 - 150 L/min
Minimum measured quantity	:	See § 1.2 of the description
	Further properties are described in the annexes: – Description T11020 revision 0; – Documentation folder T11020-1.	
Valid until	27 February 2027	

Issuing Authority **NMI Certin B.V., Notified Body number 0122**  
27 February 2017



C. Oosterman  
Head Certification Board

**NMI Certin B.V.**  
Hugo de Grootplein 1  
3314 EG Dordrecht  
The Netherlands  
T +31 78 6332332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMI Certin BV.as Notified Body can be verified at <http://ec.europa.eu/enterprise/newapproach/nando/>

Reproduction of the complete document only is permitted.

