

Issued by

NMi Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to conformity assessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer

Endress + Hauser B.V.
Nikkelstraat 6
1411 AJ Naarden
The Netherlands

Measuring instrument

An interruptible or non-interruptible **measuring system** for liquids other than water, based on a dynamic mass (Coriolis) meter.

Manufacturer's mark or name : Endress + Hauser

Type designation : MID 001

Accuracy class : 0,3; 0,5; 1,0 or 1,5

Destined for the measurement of : Oil and oil products, chemicals, potable liquids and liquefied gasses under pressure with densities between 400 and 1400 kg/m³.

Further properties are described in the annexes:

- Description T10221 revision 9;
- Documentation folder T10221-3.

Valid until

15 December 2029

Initially issued

15 December 2009

Remarks

- This revision replaces the previous version(s), except for its documentation folder.
- The measuring system is approved for measuring mass, density and volume.

Issuing Authority

NMi Certin B.V., Notified Body number 0122
23 July 2024

Certification Board

NMi Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMi Certin B.V. as Notified Body can be verified at <http://ec.europa.eu/growth/tools-databases/nando/>

Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.