

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to conformity assessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Endress+Hauser International AG
Kägenstrasse 2
4153 Reinach
Switzerland

Measuring instrument An interruptible or non-interruptible **measuring system** for liquids other than water, based on a dynamic mass (coriolis) meter
Type designation : Flow Metering System
Accuracy class : 0,3; 0,5; 1,0; 1,5; 2,5
Destined for the measurement of : See § 1.2 of the description

Further properties are described in the annexes:

- Description T10728 revision 4;
- Documentation folder T10728-4.

Valid until 1 February 2026

Remark

- This revision replaces the previous revisions.
- The documentation folder does not replace the previous documentation folder.
- The measuring system is approved for measuring mass, density and volume.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
2 May 2024

Certification Board

NMI Certin B.V.
Thijssseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMI Certin B.V. as Notified Body can be verified at <http://ec.europa.eu/growth/tools-databases/nando/>

Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.