



EC-type examination certificate

Number **T10005** revision 0
Project number 602167
Page 1 of 1

Issued by NMI Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands

Notified Body number 0122

In accordance with The Metrologiewet (Stb. 2006, 137) as Dutch implementation of Directive 2004/22/EC on measuring instruments (MID).

Manufacturer Ishida Co., Ltd.
959-1, Shimomagari
Ritto
Shiga 520-3026
Japan

In respect of A model of an **Automatic Gravimetric filling instrument**, intended to be used as an associative selection combination weigher for Net fillings
Type : CCW-R-xxx and CCW-RS-xxx

Characteristics Accuracy class X ref (1)
Electromagnetic environment class E2
Temperature range -5 °C / +40 °C

In the description number T10005 revision 0 further characteristics are described.

Valid until 8 June 2017

Description and documentation The instrument is described in the description number T10005 revision 0 and documented in the documentation folder T10005-1, appertaining to this EC-type examination certificate.

Dordrecht, 8 June 2007
NMI Certin B.V.

Ing. C. Oosterman
Manager Product Certification