

## EC-type examination certificate

Number **T10010** revision 2 Project number 811252 Page 1 of 1

Issued by NMi Certin B.V.

Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands

In accordance

with + + +

The Metrologiewet (Stb. 2006, 137) as Dutch implementation of Directive 2004/22/EC

on measuring instruments (MID).

Manufacturer Welvaarts weegsystemen

De Tweeling 4

5215 MC 's-Hertogenbosch

The Netherlands

In respect of A model of an **Automatic Catchweighing instrument**, intended to be used for

mobile weighing applications.

Type : Depending on the configuration

Characteristics Accuracy class Y(a) or Y(b)

Mechanical environment class M3 Electromagnetic environment class E3

 $Max \le 4000 \text{ kg}$ 

Min  $\geq$  20 e for class Y(a) Min  $\geq$  10 e for class Y(b)

Min ≥ 5e for garbage weighers

 $e \ge 0.2 \text{ kg}$ 

n ≤ 2000 divisions

In the description number T10010 revision 2 further characteristics are described.

Valid until 2 April 2017

Description and The instrument is described in the description number T10010 revision 2 documentation and documented in the documentation folder T10010-1, appertaining to this

EC-type examination certificate.

Remarks / This revision replace the earlier version, except for its documentation folder

The Notified Body no. 0122 NMi Certin, 22 December 2009

C. Oosterman

**Head Certification Board** 

NMi Certin BV Hugo de Grootplein 1 3314 EG Dordrecht PO Box 394 3300 AJ Dordrecht, NL T +31 78 6332332 F +31 78 6332309 certin@nmi.nl

www.nmi.nl

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see "Regulation objection and appeal against decisions of NMi")

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

Reproduction of the complete document is permitted.

