



EC-type examination certificate

Number **T10041** revision 1
Project number 607407
Page 1 of 1

Issued by NMI Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands

Notified Body number 0122

In accordance with The Metrologiewet (Stb. 2006, 137) as Dutch implementation of Directive 2004/22/EC on measuring instruments (MID).

Manufacturer Avery-Weigh-Tronix
Foundry Lane, Smetwick,
West Midlands B66 2LP
United Kingdom

In respect of A model of an **Automatic Catchweighing instrument**, intended to be used as a checkweigher, weightgrader, weigh labeler or weigh price labeler.
Type : A-3xxx

Characteristics Accuracy class XIII(1) and Y(a)
Electromagnetic environment class E1 or E2
Temperature range -10 °C / +40 °C

In the description number T10041 revision 1 further characteristics are described.

Valid until 23 November 2016

Description and documentation The instrument is described in the description number T10041 revision 1 and documented in the documentation folder T10041-1, appertaining to this EC-type examination certificate.

Remarks This revision EC type-examination certificate replaces the earlier version, except for its documentation folder.

Dordrecht, 11 December 2006
NMI Certin B.V.

Ing. C. Oosterman
Manager Product Certification

