

Nederlands Meetinstituut

EC-type examination certificate

Number T10140 revision 1 Project number 710257 Page 1 of 1

Issued by

NMi Certin B.V.

Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands

Notified Body number 0122

In accordance with

The "Metrologiewet" (Stb. 2006, 137) as Dutch implementation of Directive

2004/22/EC on measuring instruments (MID).

Manufacturer

Yamato Scale GmbH

Hanns-Martin-Schleyer-Strasse 13

D47877 Willich Germany

In respect of

A model of an Automatic Catchweighing instrument, intended to be used as a

checkweigher.

Type(s):

CSG..L and CMG..L with controller CE2000;

CSG..LW and CMG..LW with controller CE2000

Characteristics

Accuracy class (OIML R51)

XIII(1)

Mechanical environment class

not applicable

Electromagnetic environment class

F2

Temperature range

-10 °C / +40 °C, 85% humidity at 40 °C

non-condensing

In the description number T10140 revision 1 further characteristics are described.

Valid until

18 April 2018

Description and The instrument is described in the description number T10140 revision 1 documentation and documented in the documentation folder T10140-1, appertaining to this

EC-type examination certificate.

Remarks

This revision EC-type examination certificate replaces the earlier versions, except for

its documentation folder.

Dordrecht, 20 October 2008

NMi Certin B.V.

Ing/C. Oosterman

Manager Product Certification

NMi Certin BV

www.nmi.nl

Hugo de Grootplein 1 3314 EG Dordrecht PO Box 394 3300 AJ Dordrecht, NL T+31 78 6332332 F+31 78 6332309 certin@nmi.nl

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see "Regulation objection and appeal against decisions of NMi")

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

Reproduction of the complete document is permitted.

