

Issued by NMI Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands

In accordance with The "Metrologiewet" (Stb. 2006, 137) as Dutch implementation of Directive 2004/22/EC on measuring instruments (MID).

Manufacturer Precia Molen
Franse Akker 1
4824 AL Breda
The Netherlands

In respect of A model of a **Discontinuous totalizing automatic weighing instrument**
Type: ABS-...-(SCS if applicable)

Characteristics

Accuracy class (OIML R107):	0.2, 0.5, 1 and 2
Climatic environment:	-10 °C / +40 °C 85% humidity at 40 °C non-condensing
Electromagnetic environment class:	E2
Mechanical environment class:	not applicable

In the description number T10162 revision 1 further characteristics are described.

Valid until 17 September 2018

Description and documentation The instrument is described in the description number T10162 revision 1 and documented in the documentation folder T10162-2, appertaining to this EC-type examination certificate.

Remarks This revision EC type-approval certificate replaces the earlier versions, including its documentation folder.

The Notified Body no. 0122
NMI Certin, 3 June 2009


C. Oosterman
Head Certification Board