

EC type-examination certificate

Number **T10315** revision 0 Project number 10200473 Page 1 of 1

Issued by NMi Certin B.V.

Hugo de Grootplein 3314 EG Dordrecht The Netherlands

In accordance

The "Metrologiewet" (Stb. 2006, 137) as Dutch implementation of Directive

2004/22/EC on measuring instruments (MID).

Manufacturer Laurijsen Weegautomaten Dongen B.V.

De Leest 22 5107 RC Dongen The Netherlands

In respect of A model of an Automatic Gravimetric Filling instrument intended to be used as

cumulative weigher.

Type: Newweigh D-serie

Characteristics Reference accuracy class X (OIML R61) Ref(1)

Mechanical environment class M1
Electromagnetic environment class E1

Climatic environment as mentioned in the test certificates involved

In the description number T10315 revision 0 further characteristics are described.

Valid until 24 September 2020

Description and The instrument is described in the description number T10315 revision 0 documentation and documented in the documentation folder T10315-1, appertaining to this

EC-type examination certificate.

The Notified Body no. 0122 NMi Certin, 24 September 2010

C. Oosterman

Head Certification Board

NMi Certin B.V. Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands T +31 78 6332332 certin@nmi.nl www.nmi.nl This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The designation of NMi Certin BV. as Notified Body can be verified at http:// ec.europa.eu/enterprise/newapproach/nand Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see "Regulation objection and appeal against decisions of NMi" www.nmi.nl)

Reproduction of the complete document only is permitted

