

Issued by NMI Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands

In accordance with The "Metrologiewet" (Stb. 2006, 137) as Dutch implementation of Directive 2004/22/EC on measuring instruments (MID).

Manufacturer Group Pack S.r.l.
Strada del Lavoro, 71
47892, Acquaviva
Repubblica San Marino

In respect of A model of an **Automatic Gravimetric Filling instrument** intended to be used as cumulative weigher, subtractive weigher or single load weigher.
Type: TE-series

Characteristics Reference accuracy class X (OIML R61) Ref(1)
Mechanical environment class not applicable
Electromagnetic environment class E2
Temperature range +5 °C / +40 °C, 85% humidity at 40 °C
non-condensing

In the description number T10501 revision 0 further characteristics are described.

Valid until 2 July 2022

Description and documentation The instrument is described in the description number T10501 revision 0 and documented in the documentation folder T10501-1, appertaining to this EC-type examination certificate.

The Notified Body no. 0122
NMI Certin, 2 July 2012

C. Oosterman
Head Certification Board