



Issued by NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands

Notified Body number 0122

In accordance with The Metrologiewet (Stb. 2006, 137) as Dutch implementation of Directive 2004/22/EC on measuring instruments (MID).

Manufacturer SICK AG.
Nimburger Strasse 11
D-79276 Reute
Germany

In respect of A model of a **multi-dimensional measuring instrument** for measuring rectangular, non-rectangular, irregular shaped, non reflective and opaque boxes.
Type : VMS 520

Characteristics Mechanical environment class M3
Electromagnetic environment class E2
Temperature range -10 °C / +40 °C

In the description number T10019 revision 3 further characteristics are described.

Valid until 16 March 2017

Description and documentation The instrument is described in the description number T10019 revision 3 and documented in the documentation folder T10019-1, appertaining to this EC-type examination certificate.

Remarks This revision EC-type examination certificate replaces the earlier version, except for its documentation folder.

Dordrecht, 16 November 2007
NMi Certin B.V.

Ing. C. Oosterman
Manager Product Certification