

Issued by	NMi Certin B.V. Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands
	Notified Body number 0122
In accordance with	The "Metrologiewet" (Stb. 2006, 137) as Dutch implementation of Directive 2004/22/EC on measuring instruments (MID).
Manufacturer	Brain Bee S.p.A. a Socio Unico Via Quasimodo 4/a 43126 Parma Italy
In respect of	A model of an Exhaust Gas Analyser . Manufacturer's mark or name : Boston Type : AGS-688
Characteristics	Accuracy class 0 Electromagnetic environment class E2 Mechanical environment class M1 Temperature range +5 °C / +40 °C (non-condensing) In the description number T10145 revision 0 further characteristics are described.
Valid until	21 May 2018
Description and documentation	The instrument is described in the description number T10145 revision 0 and documented in the documentation folder T10145-1, appertaining to this EC-type examination certificate.
Remarks	Revision 1 was issued due to the addition of a pump. This revision replaces the earlier version, except for its documentation folder.

Issuing Authority

NMi Certin B.V., Notified Body number 0122
12 February 2010


C. Oosterman
Head of the Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The designation of NMI Certin B.V. as Notified Body can be verified at <http://ec.europa.eu/enterprise/newapproach/nando/>

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).

Reproduction of the complete document only is permitted.