

EC type-approval certificate

Number T2777 revision 3 Project number 602184 Page 1 of 4

Issued by

NMi Certin B.V.

Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands

The Council Directive 2009/2

with

Manufacturer Mettler-Toledo AG

> Im Langacher 8606 Greifensee Switzerland

In respect of

Manufacturer mark/name:

Type HR...and HG... Moisture analyzer

Characteristics

Max ≤ 81 g

e ≥ 1 mg

n ≤ 81000 divisions

Temperature range 17.5 °C / 22.5

Valid until

22 February 2016

Description and The instrument is described in the description number T2777 revision 3 and documentation documented in the documentation folder T2777-1, appertaining to this

EC type-approval certificate.

Remarks

This revision EC type-approval certificate replaces the earlier versions, except for

documentation folder.

The Notified Body no. 0122 NMi Certin, 28 January 2011

Ooste mar

NMi Certin B.V. Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands T+31 78 6332332 certin@nmi.nl www.nmi.nl

that no liability is accepted and that the applicant shall indemnify third-party liability

The designation of NMi Certin BV. as Notified Body can be verified at http://

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see "Regulation objection and appeal against decisions of NMi" www.nmi.nl)

Reproduction of the complete document only is permitted

