

Issued by NMI Certin B.V.  
Hugo de Grootplein 1  
3314 EG Dordrecht  
The Netherlands

In accordance with The Council Directive 90/384/EEC on non-automatic weighing instruments.

Applicant Mettler-Toledo GmbH  
Im Langacher  
8606 Greifensee  
Switzerland

In respect of A class **I**, electronic, single-interval **non-automatic weighing instrument**.  
Manufacturer : Mettler-Toledo  
Type : SAG

Characteristics  $n \leq 210000$  divisions  
 $\text{Max} \leq 210$  g  
 $e \geq 1$  mg  
Temperature range 20 °C / 25 °C  
In the description number T5412 revision 3 further characteristics are described.

Valid until 19 January 2019

Description and documentation The instrument is described in the description number T5412 revision 3 and documented in the documentation folder T5412-2, appertaining to this EC type-approval certificate.

Remarks This revision replaces the earlier version, except for its documentation folder.

The Notified Body, No. 0122  
NMI Certin, 7 May 2009

C. Oosterman  
Head Certification Body