

EC type-approval certificate

Number **T5520** revision 6
Project number 812019
Page 1 of 4

Issued by NMI Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands

In accordance with The Council Directive 90/384/EEC on non-automatic weighing instruments.

Applicant ARJO Huntleigh
Medical Beds Division
Unit 3 Britannia Park
Trident Drive
Wednesbury, WS10 7XB
United Kingdom

In respect of A class **(III)** or **(III)**, electronic, single-interval **non-automatic weighing instrument**.
Manufacturer : Huntleigh
Type : Contoura / C980B

Characteristics $n \leq 1800$ divisions
 $180 \text{ kg} \leq \text{Max} \leq 500 \text{ kg}$ or $400 \text{ lb} \leq \text{Max} \leq 1100 \text{ lb}$
 $e \geq 100 \text{ g}$ or $e \geq 0.5 \text{ lb}$
In the description number T5520 revision 6 further characteristics are described.

Valid until 20 May, 2019

Description and documentation The instrument is described in the description number T5520 revision 6 and documented in the documentation folder T5520-3, appertaining to this EC type-approval certificate.

Remarks This revision replaces the earlier versions, except for its documentation folder.

The Notified Body, No. 0122
NMI Certin, 20 May 2009


C. Oosterman
Head Certification Body