

Issued by	NMi Certin B.V. Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands
In accordance with	The Council Directive 2009/23/EC on non-automatic weighing instruments.
Manufacturer	Mettler-Toledo 1150 Dearborn Drive Worthington, Ohio 43085 United States of America
In respect of	A class (III) or (III) , electronic, single- or multi-range non-automatic weighing instrument . Manufacturer mark/name: Mettler-Toledo Type : 8270
Characteristics	$2 \text{ kg} \leq \text{Max} \leq 20 \text{ kg}$ $e \geq 0.002 \text{ kg}$ $n \leq 4000$ divisions (per partial weighing range with a maximum of 2 partial weighing ranges for class (III) instruments or $n \leq 1000$ divisions for class (III) instruments Temperature range $0 \text{ }^\circ\text{C} / 40 \text{ }^\circ\text{C}$ In the description number T5530 revision 2 further characteristics are described.
Valid until	9 August 2019
Description and documentation	The instrument is described in the description number T5530 revision 2 and documented in the documentation folder T5530-2, appertaining to this EC type-approval certificate.
Remarks	This revision EC type-approval certificate replaces the earlier versions, except for its documentation folder.

The Notified Body no. 0122
NMi Certin, 7 September 2011



C. Oosterman
Head Certification Board