

## EC type-approval certificate

Number **T5564** revision 3 Project number 812334 Page 1 of 5

Issued by

NMi Certin B.V

Hugo de Grootplein 3314 EG Dordrecht The Netherlands

In accordance

The Council Directive 90/384/EEC on non-automatic weighing instruments.

with

Applicant DIBAL S.A.

Astintze Kalea, 24-Pol. Ind. Neinver 48016 Derio (Bilbao – Vizcaya)

SPAIN

In respect of

A class (III), electronic, single-interval non-automatic weighing instrument,

(intended to be used for direct sales to the public).

Manufacturer

DIBAL

Type

KIT COMPACT, KIT GLOBAL or KIT GLOBAL PLUS

Characteristics

 $n \leq 3000 \ divisions$ 

 $6 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$ 

e ≥ 2 g

Temperature range -10 °C / 40 °C

In the description number T5564 revision 3 further characteristics are described.

Valid until

31 July 2019

Description and

The instrument is described in the description number T5564 revision 3 and documented in the documentation folder T5564-1, appertaining to this EC

type-approval certificate.

Remarks

This revision EC type-approval certificate replaces the earlier versions, except for its

documentation folder.

The Notified Body No. 0122 NMi Certin, 2 November 2009

C. Øosterman

**Head Certification Board** 

NMi Certin BV Hugo de Grootplein 1 3314 EG Dordrecht PO Box 394 3300 AJ Dordrecht, NL T+31 78 6332332 F+31 78 6332309

certin@nmi.nl

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see "Regulation objection and appeal against decisions of NMi") This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

Reproduction of the complete document is permitted.

