

Issued by NMI Certin B.V.
 Hugo de Grootplein 1
 3314 EG Dordrecht
 The Netherlands

In accordance with The Council Directive 90/384/EEC on non-automatic weighing instruments.

Manufacturer Mettler-Toledo (Albstadt) GmbH
 Unter dem Malesfelsen 34
 D-72458 Albstadt
 Germany

In respect of A class **II** or **III**, electronic **non-automatic weighing instrument**
 Manufacturer: Mettler-Toledo Ohaus Testut
 Type : VIPER ... Ranger... Cruiser...
 BBK 4...

Characteristics

II	III multi-interval
n ≤ 61000 divisions	n ≤ 10000 divisions (per weighing range)
610 g ≤ Max ≤ 35.1 kg	1.5 kg ≤ Max ≤ 35.1 kg For the additional load receptor 1.5 kg ≤ Max ≤ 6000 kg
e ≥ 10 mg	e ≥ 1 g
e = d or e = 10d	e = d
+10 °C / +30 °C	+5 °C / +35 °C


In the description number T5627 revision 8 further characteristics are described.

Valid until 4 February 2020

Description and documentation The instrument is described in the description number T5627 revision 8 and documented in the documentation folder T5627-7, appertaining to this EC type-approval certificate.

Remarks This revision replaces the earlier versions, except for its documentation folder.

The Notified Body No. 0122
 NMI Certin, 24 February 2010



C. Oosterman
 Head Certification Board

NMI Certin BV
 Hugo de Grootplein 1
 3314 EG Dordrecht
 PO Box 394
 3300 AJ Dordrecht, NL
 T +31 78 6332332
 F +31 78 6332309
 certin@nmi.nl
 www.nmi.nl

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMI (see "Regulation objection and appeal against decisions of NMI")

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

Reproduction of the complete document is permitted.