

type-approval certificate

Number T5744 revision 6 Project number 10200404 Page 1 of 5

Issued by

NMi Certin B.V.

Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands

In accordance

with

Manufacturer Ishida Co., Ltd.

44 Sanno-cho, Shogoin,

Sakyo-ku, Kyoto-city, 606-8392

Japan

In respect of

A class (III), electronic, single-interval **non-automatic we**i

intended to be used for direct sales to the public. Ishida Co., Ltd.;

Manufacturer mark/name

MICRON S.r.I.

ASTRA

Characteristics

6 kg ≤ Max ≤ 15 kg	6 kg ≤ Max ≤ 15 kg
n ≤ 3000 divisions	n ≤ 6000 divisions
e ≥ 2 g	e ≥ 2 g
-5 °C / +40 °C	+5 °C / +35 °C
T ≤ 67% of Max	T ≤ 84% of Max

In the description number T5744 revision 6 further characteristics are described.

Valid until

30 June 2020

Description and The instrument is described in the description number T5744 revision 6 and documentation documented in the documentation folder T5744-4, appertaining to this

EC type-approval certificate.

Remarks

This revision replaces the earlier version, except for its documentation folder.

The Notified Body no. 0122 NMi Certin, 26 July 2010

C. Oosterman

Head Certification Board

NMi Certin B.V. Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands T+31 78 6332332 certin@nmi.nl www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The designation of NMi Certin BV. as Notified Body can be verified at http://

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see Regulation objection and appeal against decisions of NMi" www.nmi.nl)

Reproduction of the complete document only is permitted

