

## type-approval certificate

Number T5751 revision 2 Project number 10200897 Page 1 of 5

Issued by

NMi Certin B.V.

Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands

In accordance

The Council Directive 2009/23/

with

**Shekel Electronic Scales** Manufacturer

Kibbutz Beit Keshet M.P. Lower Galilee 15247 ISRAEL

In respect of

A class (III) graduated, self-indicating, electronic, single-interval, non-automatic

weighing instrument, intended to be used for direct sales to the public.

Manufacturer mark/name: Shekel Electronic Scales SC-900 Type

Characteristics

 $6 \text{ kg} \leq \text{Max} \leq 15 \text{ kg}$ 

 $e \ge 5q$ 

n ≤ 3000 divisions

Temperature range +10 °C / +40 °C.

In the description number T5751 revision 2 further characteristics are

Valid until

7 August 2020

Description and The instrument is described in the description number T5751 revision 2 and documentation documented in the documentation folder T5751-2, appertaining to this

EC type-approval certificate.

Remarks

This revision EC type-approval certificate replaces the earlier versions, except for

documentation folder.

The Notified Body no. 0122 NMi Centin, 7-August 2010

**Oostermar** 

Head Certification Board

NMi Certin B.V. Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands T+31 78 6332332 certin@nmi.nl www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The designation of NMi Certin BV. as Notified Body can be verified at http:// ec.europa.eu/enterprise/newapproach/nando/

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see "Regulation objection and appeal against decisions of NMi" www.nmi.nl)

Reproduction of the complete document only is permitted

