


Issued by	NMi Certin B.V. Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands
In accordance with	The Council Directive 2009/23/EC on non-automatic weighing instruments.
Manufacturer	TERAOKA Weigh-System Pte. Ltd. 4 Leng Kee Road, #06-01 SIS Building SINGAPORE 159088
In respect of	A class (III) , electronic, single- or multi-interval, non-automatic weighing instrument , intended to be used for direct sales to the public. Manufacturer : Teraoka or P.T. Tropical Electronic Type : SM-700..
Characteristics	$n \leq 3000$ divisions (per partial weighing range) $n \leq 7500$ divisions (single range) $3 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$ $e \geq 1 \text{ g}$ maximum of two partial weighing ranges $T \leq -\text{Max}_1$ In the description number T6017 revision 8 further characteristics are described.
Valid until	28 February 2022
Description and documentation	The instrument is described in the description number T6017 revision 8 and documented in the documentation folder T6017-4, appertaining to this EC type-approval certificate.
Remarks	This revision replaces the earlier versions, except for its documentation folder.

The Notified Body no. 0122
NMi Certin, 28 February 2012



C. Oosterman
Head Certification Board