

EC type-approval certificate

Number T6527 revision 3 Project number 811878 Page 1 of 4

Issued by NMi Certin B.V.

> Hugo de Grootplein 3314 EG Dordrecht The Netherlands

In accordance

with

The Council Directive 90/384/EEC on

Applicant

OHAUS Corporation 19A Chapin Road Pine Brook, NJ 07058 **United States of America**

In respect of

A class (III), electronic, single-interval, non-automatic weighing instrument. **OHAUS**

Manufacturer

FD Series

Characteristics

 $3 \text{ kg} \leq \text{Max} \leq 15 \text{ kg}$

e ≥ 1 g

n ≤ 3000 divisions

Temperature range 0 °C / +40 °C

In the description number T6527 revision 3 fur

5 July 2014

Description and The instrument is described in the description number T6527 revision 3 and documentation documented in the documentation folder T6527-3, appertaining to this

EC type-approval certificate.

Remarks

This revision EC type-approval certificate replaces the earlier versions, including its

documentation folder.

The Notified Body no. 0122 NMi Certin 5 March 2009

C. Oøsterman

NMi Certin BV Hugo de Grootplein 1 3314 EG Dordrecht PO Box 394 3300 AJ Dordrecht, NL T+31 78 6332332 F+31 78 6332309 certin@nmi.nl www.nmi.nl

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see "Regulation objection This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

Reproduction of the complete document is

