

EC type-approval certificate

Number **T7228** revision 2 Project number 811878 Page 1 of 5

Issued by NMi Certin B.V.

Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands

In accordance

with

The Council Directive 90/384/EEC on non-automatic weighing instruments.

Applicant + OHAUS Corporation

19A Chapin Road Pine Brook, NJ 07058 United States of America

In respect of A class (III), electronic, single-interval, non-automatic weighing instrument.

Manufacturer : OHAUS

Type : Valor 5000 series / V51Pxx

Characteristics $3 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$

e ≥ 1 g

 $n \le 3000$ divisions

Temperature range 0 °C / +40 °C

In the description number T7228 revision 2 further characteristics are described.

Valid until 9 August 2017

Description and The instrument is described in the description number T7228 revision 2 and documentation documented in the documentation folder T7228-2, appertaining to this

EC type-approval certificate.

Remarks This revision EC type-approval certificate replaces the earlier versions, including its

documentation folder.

The Notified Body no. 0122 NMi Certin 5 March 2009

C. Oosterman

Head Certification Board

NMi Certin BV
Hugo de Grootplein 1
3314 EG Dordrecht
PO Box 394
3300 AJ Dordrecht, NL
T +31 78 6332332
F +31 78 6332309
certin@nmi.nl

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see "Regulation objection and appeal against decisions of NMi")

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

Reproduction of the complete document is permitted.

