

Issued by	NMi Certin B.V. Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands
In accordance with	The Council Directive 2009/23/EC on non-automatic weighing instruments.
Manufacturer	Libra Engineering Ltd. Lubljana Str. 46 1618 Sofia Bulgaria
In respect of	A class <b>(III)</b> , electronic, single- or multi-interval, <b>non-automatic weighing instrument</b> , intended to be used for direct sales to the public. Manufacturer mark/name: Libra Engineering Type : LE - PCS, LE - PCSC or LE - WS
Characteristics	$3 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$ $e \geq 1 \text{ g}$ $n \leq 3000$ divisions (per partial weighing range) Maximum of two partial weighing ranges $T \leq -\text{Max}$ Temperature range $-10^\circ\text{C} / +40^\circ\text{C}$  In the description number T7333 revision 2 further characteristics are described.
Valid until	12 December 2017
Description and documentation	The instrument is described in the description number T7333 revision 2 and documented in the documentation folder T7333-3, appertaining to this EC type-approval certificate.
Remarks	This revision EC type-approval certificate replaces the earlier versions, including its documentation folder.

The Notified Body no. 0122  
NMi Certin, 7 February 2011



C. Oosterman  
Head Certification Board

**NMi Certin B.V.**  
Hugo de Grootplein 1  
3314 EG Dordrecht  
The Netherlands  
T +31 78 6332332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The designation of NMi Certin BV. as Notified Body can be verified at <http://ec.europa.eu/enterprise/newapproach/nando/>

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see "Regulation objection and appeal against decisions of NMi" [www.nmi.nl](http://www.nmi.nl))

Reproduction of the complete document only is permitted