

Issued by	NMi Certin B.V. Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands
In accordance with	The Council Directive 2009/23/EC on non-automatic weighing instruments.
Applicant	Micron S.R.L. Via Ticino 53 21044 Cavarina (VA) Italy
In respect of	A class (III) , electronic, single- or multi-interval, non-automatic weighing instrument , intended to be used for direct sales to the public. Manufacturer : Micron Type : ASP / ATP / AHW / AHC / QSP / QTP / QHW / QHC
Characteristics	$3 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$ $e \geq 1 \text{ g}$ $n \leq 3000$ divisions (per partial weighing range) Maximum of 2 partial weighing ranges $T \leq -\text{Max}$ Temperature range $-10^{\circ}\text{C} / +40^{\circ}\text{C}$ In the description number T7415 revision 1 further characteristics are described.
Valid until	20 May 2018
Description and documentation	The instrument is described in the description number T7415 revision 1 and documented in the documentation folder T7415-2, appertaining to this EC type-approval certificate.
Remarks	This revision EC type-approval certificate replaces the earlier version, including its documentation folder.

Notified Body Number 0122
NMi Certin, 10 November 2010



C. Oosterman
Head Certification Board