

EC type-approval certificate

Number **T7592** revision 1 Project number 9200170 Page 1 of 5

Issued by NMi Certin B.V

Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands

In accordance

The Council Directive 90/384/EEC on non-automatic weighing instruments.

with

Applicant Inwag Systems Renata Wlodarczyk

ul. Legionow Polskich 43 99-220 Wartkowice

Poland

In respect of A class (III), electronic, single- or multi-interval, non-automatic weighing

instrument, intended to be used for direct sales to the public.

Manufacturer : Inwag

Type : Aspen / Aspen T / Aspen H

Queen / Queen T / Queen H

Characteristics $3 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$

e ≥ 1 g

n ≤ 3000 divisions (per partial weighing range)

Maximum of 2 partial weighing ranges

T ≤ -Max

Temperature range -10°C / +40°C

In the description number T7592 revision 1 further characteristics are described.

Valid until + 25 June 2019

Description and The instrument is described in the description number T7592 revision 1 and

documentation documented in the documentation folder T7592-1, appertaining to this

EC type-approval certificate.

Remarks This revision replaces the earlier version, except for its documentation folder.

The Notified Body No. 0122 NMi Certin, 25 June 2009

C. Øosterman

Head Certification Board

NMi Certin BV
Hugo de Grootplein 1
3314 EG Dordrecht
PO Box 394
3300 AJ Dordrecht, NL
T +31 78 6332332
F +31 78 6332309
certin@nmi.nl

www.nmi.nl

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see "Regulation objection and appeal against decisions of NMi") This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

Reproduction of the complete document is permitted.

