

EC type-approval certificate

Number T7675 revision 0 Project number 9200879 Page 1 of 5

Issued by

NMi Certin B.V.

Hugo de Grootplein 3314 EG Dordrecht The Netherlands

In accordance

with

Manufacture Dini Argeo S.r.l.

Via della Fisica n. 20

41042 Spezzano di Fiorano, Modena

Italy

In respect of

electronic, single- or multi-interval non-automatic weighing

instrument.

Manufacturer

Dini Argeo S.r.l.

Type ASG / ASGP / ASGPP (direct sales to the public)

ASGW / ASGWP (not for direct sales to the

 $3 \text{ kg} \leq \text{Max} \leq 45 \text{ kg}$

e ≥ 1 g

n ≤ 3000 divisions (per partial weighing range)

Maximum of 2 partial weighing ranges

 $T \leq -Max$

Temperature range -10 °C / +40 °C

In the description number T7675 revision 0 further characteristics are described

Valid until

11 June 2020

Description and The instrument is described in the description number T7675 revision 0 and documentation documented in the documentation folder T7675-1, appertaining

EC type-approval certificate

The Notified Body no. 0122 NMi Centin, 11 June 2010

Certification Board

Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands T+31 78 6332332 certin@nmi.nl www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The designation of NMi Certin BV. Body can be verified at http://

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see "Regulation objection and appeal against decisions of NMi" www.nmi.nl)

Reproduction of the complete document only is permitted

