

## EC type-approval certificate

Number T7684 revision 0 Project number 9200780 Page 1 of 5

Issued by

NMi Certin B.V.

Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands

In accordance

with

The Council Directive 2009/23/EC

Manufacturer

Mettler-Toledo (Changzhou) Measurement Technology

111 West Taihu Road XinBei District, Changzhou

Jiangsu 213125

Peoples Republic of China

In respect of

A class (III), electronic, single-interval, non-automatic we

Manufacturer mark/name:

Mettler-Toledo

PCA320, PCA220 series

Characteristics

 $100 \text{ kg} \le \text{Max} \le 10000 \text{ kg}$ 

e ≥ 50 q

 $n \le 3000$  divisions

Temperature range -10 °C / +40 °C

In the description number T7684 revision

Valid until

25 October 2020

Description and The instrument is described in the description number T7684 revision 0 and documentation documented in the documentation folder T7684-1, appertaining to this

EC type-approval certificate.

The Notified Body no. 0122 NMi Centin, 25 October 2010

Head Certification Board

NMi Certin B.V. Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands T+31 78 6332332 certin@nmi.nl www.nmi.nl

that no liability is accepted and that the applicant shall indemnify third-party liability.

The designation of NMi Certin BV. as Notified

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see 'Regulation objection and appeal against decisions of NMi" www.nmi.nl)

document only is permitted

