

Issued by NMI Certin B.V.
 Hugo de Grootplein 1
 3314 EG Dordrecht
 The Netherlands

In accordance with The Council Directive 2009/23/EC on non-automatic weighing instruments.

Manufacturer MeWa Schwering GmbH
 Nikolaus-Otto-Straße 5a
 D-19061 Schwerin
 Germany

In respect of A class **I** or **II**, electronic, self-indicating **non-automatic weighing instrument**.
 Manufacturer : MeWa Schwerin
 Type : PBD/PDC

Characteristics

Class I	Class II
$n \leq 82000$ divisions	$n \leq 42000$ divisions
$500 \text{ g} \leq \text{Max} \leq 6200 \text{ g}$ or $2500 \text{ ct} \leq \text{Max} \leq 31000 \text{ ct}$	$50 \text{ g} \leq \text{Max} \leq 12 \text{ kg}$ or $250 \text{ ct} \leq \text{Max} \leq 60000 \text{ ct}$
$e \geq 0,01 \text{ g}$ or $e \geq 0,1 \text{ ct}$	$e \geq 0,01 \text{ g}$ or $e \geq 0,1 \text{ ct}$
$e = d$ or $e = 10d$	$e = d$ or $e = 10d$

Temperature range +10 °C / +30 °C
 In the description number T7697 revision 1 further characteristics are described.

Valid until 4 February 2020

Description and documentation The instrument is described in the description number T7697 revision 1 and documented in the documentation folder T7697-2, appertaining to this EC type-approval certificate.

Remarks This revision EC type-approval certificate replaces the earlier version, including its documentation folder.

The Notified Body no. 0122
 NMI Certin, 17 June 2011

C. Oosterman
 Head Certification Board

NMI Certin B.V.
 Hugo de Grootplein 1
 3314 EG Dordrecht
 The Netherlands
 T +31 78 6332332
 certin@nmi.nl
 www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The designation of NMI Certin BV. as Notified Body can be verified at <http://ec.europa.eu/enterprise/newapproach/nando/>

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMI (see "Regulation objection and appeal against decisions of NMI" www.nmi.nl)

Reproduction of the complete document only is permitted