

| | |
|-------------------------------|--|
| Issued by | NMi Certin B.V. Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands |
| In accordance with | The Council Directive 2009/23/EC on non-automatic weighing instruments. |
| Manufacturer | R. Slonskio Enterprise "Romosas" Vilniaus 121 LT-76354, Siauliai Luthuania |
| In respect of | A class (III) , electronic, single- or multi-interval non-automatic weighing instrument . |
| Manufacturer mark/name: | R. Slonskio Enterprise "Romosas" |
| Type | MR / MRP / MPP (direct sales to the public) MW / MWP (not for direct sales to the public) |
| Characteristics | $3 \leq \text{Max} \leq 45 \text{ kg}$ $e \geq 1 \text{ g}$ $n \leq 3000$ divisions (per partial weighing range) Maximum of two partial weighing ranges $T \leq -\text{Max}$ Temperature range $-10 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$ In the description number T7709 revision 1 further characteristics are described. |
| Valid until | 4 May 2020 |
| Description and documentation | The instrument is described in the description number T7709 revision 0 and documented in the documentation folder T7709-2, appertaining to this EC type-approval certificate. |
| Remarks | This revision EC type-approval certificate replaces the earlier version, including its documentation folder. |

The Notified Body no. 0122
NMi Certin, 18 March 2011



C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The designation of NMi Certin BV. as Notified Body can be verified at <http://ec.europa.eu/enterprise/newapproach/nando/>

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see "Regulation objection and appeal against decisions of NMi" www.nmi.nl)

Reproduction of the complete document only is permitted