

## EC type-approval certificate

Number T7730 revision 0 Project number 10200444 Page 1 of 5

Issued by

NMi Certin B.V.

Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands

In accordance

with

MeWa GmbH Manufacturer

> Nikolaus-Otto-Stasse D-19061 Schwerin

Germany

In respect of

A class (III), electronic, single- or multi-interval, multi-range **non-automatic** weighing instrument, intended to be used for direct sales to the public or not for direct sales to the public.

Manufacturer mark/name: MeWa

: LWX100, LWX100T, LWD100T (direct sales to the pu

: ZW60, ZW70 (not for direct sales to the public)

 $3 \text{ kg} \leq \text{Max} \leq 45 \text{ kg}$ 

e ≥ 1 g

 $n \le 3000$  divisions (per partial weighing range)

Maximum of 2 partial weighing ranges

T ≤ -Max

Temperature range -10°C / +40°C

In the description number T7730 revision 0 further characteristics are described

Valid until

20 May 2020

Description and The instrument is described in the description number T7730 revision 0 documentation documented in the documentation folder T7730-1, appertaining to this

EC type-approval certificate.

The Notified Body no. 01 NMi Certin, 20 May 2010

Head Certification Board

Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands T+31 78 6332332 certin@nmi.nl www.nmi.nl

that no liability is accepted and that the

The designation of NMi Certin BV. as Notified Body can be verified at http://

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see Regulation objection and appeal against decisions of NMi" www.nmi.nl)

Reproduction of the complete document only is permitted

