

## EC type-approval certificate

Number T7859 revision 0 Project number 10201077 Page 1 of 4

Issued by

NMi Certin B.V.

Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands

In accordance

with

Plentznerstrasse 16b A-4810 Gmunden

Austria

In respect of

A class (III), electronic, single- or multi-interval, non-automatic weighing

instrument, intended to be used for direct sales to the public.

Manufacturer mark/name:

ASM Sautner GmbH

Type

ASM ES-100

 $1,5 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$ 

 $e \ge 0.5 q$ 

 $n \le 7500$  divisions for single-interval instruments n ≤ 3000 divisions for multi-interval instruments

Maximum of 2 partial weighing ranges

T ≤ -Max for single interval instruments or -Max₁ for multi interval instr

Temperature range -10 °C / +40 °C

In the description number T7859 revision 0 further characteristics are described.

Valid until

20 December 2020

Description and The instrument is described in the description number T7859 revision 0 and documentation documented in the documentation folder T7859-1, appertaining

EC type-approval certificate

The Notified Body no. 0122 NMi Certin, 20 December 2010

Øoster

Head Certification Board

NMi Certin B.V. Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands T+31 78 6332332 certin@nmi.nl www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The designation of NMi Certin BV. as Notified Body can be verified at http://

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see Regulation objection and appeal

Reproduction of the complete

