

EC type-approval certificate

Number **T7860** revision 0 Project number 10201077 Page 1 of 4

Issued by

NMi Certin B.V.

Hugo de Grootplein 3314 EG Dordrecht The Netherlands

In accordance

The Council Directive 2009/23/EC on non-automatic weighing instruments.

with

Manufacturer ASM Sautner GmbH

Plentznerstrasse 16b A-4810 Gmunden

Austria

In respect of

A class (III) electronic, single- or multi-interval, non-automatic weighing

instrument.

Manufacturer mark/name:

ASM Sautner GmbH

Type

ASM ES-200

Characteristic

 $1,5 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$

e ≥ 0,5 g

 $n \leq 7500$ divisions for single-interval instruments $n \leq 3000$ divisions for multi-interval instruments

Maximum of 2 partial weighing ranges

T ≤ -Max for single interval instruments or -Max₁ for multi interval instruments

Temperature range -10 °C / +40 °C

In the description number T7860 revision 0 further characteristics are described.

Valid until

20 December 2020

Description and The instrument is described in the description number T7860 revision 0 and documentation documented in the documentation folder T7860-1, appertaining to this

EC type-approval certificate.

The Notified Body no. 0122 NMi Certin, 20 December 2010

C. Oosterman

Head Certification Board

NMi Certin B.V. Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands T +31 78 6332332 certin@nmi.nl www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The designation of NMi Certin BV. as Notified Body can be verified at http://ec.europa.eu/enterprise/newapproach/nando/

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see "Regulation objection and appeal against decisions of NMi" www.nmi.nl)

Reproduction of the complete document only is permitted

