

EC type-approval certificate

Number T7872 revision 1 Project number 10200677 Page 1 of 5

Issued by

NMi Certin B.V.

Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands

In accordance

The Council Directive 2009/23/EC on non-automatic weighing instruments.

with

Shanghai Teraoka Electronic Co., Ltd Manufacturer

Tinglin Industry Developmental Zone

Jin Shan District Shanghai 201505

Peoples Republic of China

In respect of

A class (III), electronic, single- or multi-interval non-automatic weighing

instrument, intended to be used for direct sales to the public.

Manufacturer mark/name:

Teraoka

Туре

DS-520III / DS-530III

Characteristics

 $3 \text{ kg} \leq \text{Max} \leq 600 \text{ kg}$

e ≥ 1 g

 $n \le 3000$ divisions (per partial weighing range)

Maximum of 2 partial weighing ranges T ≤ -1/2 Max for single-interval instruments T ≤ -(Max1-e1) for multi-interval instruments

Temperature range -10 °C / +40 °C

In the description number T7872 revision 1 further characteristics are described.

Valid until

19 January 2021

Description and The instrument is described in the description number T7872 revision 1 and documentation documented in the documentation folder T7872-2, appertaining to this

EC type-approval certificate.

Remark

This revision replaces the earlier version, including its documentation folder.

The Notified Body no. 0122 NMi Certin, 10 February 2011

Head Certification Board

NMi Certin B.V. Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands T+31 78 6332332 certin@nmi.nl www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The designation of NMi Certin BV. as Notified Body can be verified at http://

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see "Regulation objection and appeal against decisions of NMi" www.nmi.nl)

Reproduction of the complete document only is permitted

