

## type-approval certificate

Number T7888 revision 0 Project number 11200036 Page 1 of 4

Issued by

NMi Certin B.V.

Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands

In accordance

Manufacturer

with

**UAB Mingeda** 

Krokuvos Street 11 LT-09314 Vilnius

Lithuania

In respect of

A class (III), electronic, single- or multi-interval non-automatic weighing

instrument, intended to be used for direct sales to the public.

Manufacturer mark/name:

Mingeda

Type

MS-QTP / MS-QSP

Characteristics

 $3 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$ 

 $e \ge 1q$ 

 $n \le 3000$  divisions (per partial weighing range)

Maximum of 2 partial weighing ranges

 $T \leq -Max$ 

Temperature range -10°C / +40°C

In the description number T7888 revision 0 further characteristics are described

Valid until

24 February 2021

Description and The instrument is described in the description number T7888 revision 0 and documentation documented in the documentation folder T7888-1, appertaining to this

EC type-approval certificate.

The Notified Body no. 0122 NMi Certin, 24 February 2011

Oosterman

**Head Certification Board**